

User Guide

PROMPT Flex Caesarean Section Module

Product No: 80103



For more skills training products visit limbsandthings.com

Limbs & Things Ltd.

Sussex Street, St Philips Bristol, BS2 0RA, UK sales@limbsandthings.com +44 (0) 117 311 0500

PROMPT Flex Caesarean Section Module

Product No: 80103

This product is ideal for practising the skills required when performing a transverse incision, (both Pfannenstiel and Joel-Cohen), allowing for the layers to be sutured as well as in-situ repair of the uterine wall. Also delivering a difficult baby with an impacted head or a transverse lie can be performed.

Skills

- Abdominal and uterine wall opening and closure
- Cephalic and breech delivery
- C-section at full dilatation:
 - Head disimpaction using vaginal push methods
 - Use of balloon disimpacting device
 - Reverse breech extraction
- Transverse lie
- Instrument assisted delivery
- Delivery of the placenta
- Teamwork and human factors training in emergency Caesarean section

Features

- Uterus holds baby firmly in position for normal and transverse lies, breech and cephalic presentations
- Uterus support positions baby and directs it towards the pelvis/abdominal incision when fundal pressure is applied
- Pre-incised skin option allows easy repeat delivery practice
- Surgical pad allows realistic dissection and closure and is reversible to allow 2 uses
- Single/double in situ uterine wall closure
- Suturing of rectus sheath, fat and subcuticular or cutaneous skin closure
- Accommodates retraction using metal, plastic or rubber retractors

Package supplied

- 80104 Caesarean Section Surgical Pad (x2)
- 80138 Perineum for PROMPT Flex CDE and C-Section Modules
- 80170 Uterus for PROMPT Flex Caesarean Section Module
- 80171 Pre-incised Delivery Only Skin for PROMPT Flex Caesarean Section Module
- 80172 Surgical Skin for PROMPT Flex Caesarean Section Module
- 80173 Cervix for PROMPT Flex Caesarean Section Module
- 80175 Foam Uterus Support for PROMPT Flex Caesarean Section Module
- 80176 Full Dilatation Pelvic Insert for PROMPT Flex Caesarean Section Module



Components



Caesarean Section Surgical Pad (x2) Part No: 80104



Perineum for PROMPT Flex CDE and C-Section Modules Part No: 80138



Uterus for PROMPT Flex
Caesarean Section Module
Part No: 80170



Pre-incised Delivery Only Skin

Part No: 80171



Surgical Skin
Part No: 80172



Cervix for PROMPT Flex Caesarean Section Module

Part No: **80173**



Foam Uterus Support Part No: 80175



Full Dilatation Pelvic Insert


Part No: 80176

Attaching the C-Section Perineum

This module is to be used in conjunction with your PROMPT Flex Birthing Simulator




Remove PROMPT Flex Abdominal Skin from the PROMPT Flex Unit






Remove the Pelvic Ring Clamp from the PROMPT Flex Unit



Remove the Cervix if present



Remove the birth canal and perineum by firstly detaching the perineal skin from the poppers on the mother unit, then carefully feed the birth canal through the pelvis



Once the top half of the perineum has been detached from the unit, detach the skin from the poppers on the bottom of the mother unit




For C section at full dilation/head impaction, fit the Full Dilation Pelvic Insert in the pelvis prior to fitting the birth canal and perineum



The Full Dilation Insert allows presentation of the head just below the ischial spines, for Caesarean section at full dilation



Attach the Perineum for PROMPT Flex CDE and C-Section modules. Push the birth canal up through the pelvis





Fix the perineum top and bottom using the poppers



The perineum is now in place



If NOT simulating full dilation/head disimpaction, fit the c section cervix at this point, lining up the locating features on the flanges of the cervix and the birth canal



The cervix is now in place

Attaching the Uterus and Foam Support



To fix the uterus in place. Slide the Caesarean section uterus ring clamp into position on the clamping surface, over the birth canal (and cervix if used) flange



Rotate the securing bar into position over the uterus clamp to fix the uterus in place



Slide the Uterus Support under the uterus, ensuring the locking lug on the bottom fits into the handle recess on the PROMPT Flex mother unit



The primary function of the Uterus Support is to position the baby, and direct it towards the pelvis and abdominal incision when fundal pressure is applied


Loading the Baby into the Unit



The baby can be either loaded into position when the abdominal skin is in place, by inserting it between the skin and uterus (as shown in above photo) or it can be placed in the uterus before skin is attached. This is a matter of personal preference



Lubricate the inside of the uterus



Lubricate the cervix



Lubricate the baby. If simulating breech, reverse breech extraction or transverse lie, avoid lubricating the baby's legs



Push the baby into the uterus, through the gap between the uterus and the skin at the back of the model. Push the presenting part into the pelvis where it will come to rest against the cervix if present, or ischial spines if set up for full dilation C section



The baby can also be placed in the transverse lie position



If using the **Pre-Incised Delivery Only Skin**, attach the skin to the securing pegs on the sides and front of mother unit

Attaching the C-Section Surgical Pad to the Surgical Skin


For skills focussing on abdominal and uterine opening and closure the surgical abdomen and surgical pad are to be used



Place the Surgical Skin face down on a table



Ensure the C section Pad is the correct orientation, with the slits at the side of the pad, and the holes at the top and bottom of the pad



Pull the black straps with poppers on either side of the Surgical Skin through the slits located on the sides of the Surgical Pad



Fold the strap over and pop into place to secure



The sides of the pad are now secure



Repeat this process with the straps at the top of the Surgical Pad, ensuring that the holes of the pad align with the poppers. Pop into place to secure



Lubricate inner surface of the Caesarean Section Pad. Lightly talc the external surface of pad if sticky The Surgical Skin is now ready to be fitted to the PROMPT Flex mother unit

Attaching the Surgical Skin to the Mother Unit



Detach the perineum skin from the poppers (only at the front of the mother unit)



Attach the 2 small holes on either side of the Surgical Pad to the securing pegs on the front of the mother unit



Align the two central holes of the Surgical Pad to the two central poppers




Re-attach the perineum to the poppers




Attach the surgical skin to the abdominal securing the pegs on the sides and front of mother unit and adjust the skin accordingly to ensure a neat fit



The skin is now attached and ready for use




Tightening Tension Belt to Simulate Head Impaction




With the baby loaded into the uterus, hold the uterus firmly around the baby with one hand



With the clamp on the tensioning belt open, tighten the tensioning belt to the required degree of foetal head impaction



Close the clamp on the tensioning belt



Re attach the skin

You may also be interested in...

PROMPT Flex Dilatation & Effacement Module

Product No: 80102





To find out more visit:

limbsandthings.com



Part No: 165-007 Issue 2, March 2016 © 2016 Limbs & Things